


=====
Acq. Operator : Seq. Line : 2
Acq. Instrument : Instrument 1 Location : Vial 16
Injection Date : 7/30/2018 2:40:10 PM Inj : 1
Inj Volume : 3.0 µl
Different Inj Volume from Sequence ! Actual Inj Volume : 2.0 µl
Acq. Method : D:
Last changed : 6/6/2018 1:09:46 PM
Analysis Method : D:\CHEM32\1\1.M
Last changed : 7/30/2018 4:10:39 PM
(modified after loading)
Sample Info : A = ACN, B = 0.1% H3PO4/H2O] 0-95% A/B 14 min, 95% A/B 6 min; Athena C18 (3 µm
, 2.1 mm * 100 mm); Flow: 0.40 ml/min, T = 25

Additional Info : Peak(s) manually integrated



=====
Area Percent Report
=====

Sorted By : Signal
Multiplier : 1.0000
Dilution : 1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: DAD1 A, Sig=252,4 Ref=400,100

Peak #	RetTime [min]	Type	Width [min]	Area [mAU*s]	Height [mAU]	Area %
1	7.374	BB	0.0788	9.40683	1.42903	0.0813
2	9.972	BB	0.1808	1.15570e4	918.29095	99.9187
Totals :				1.15664e4	919.71998	

=====
*** End of Report ***